

Certificate

PROFIBUS Nutzerorganisation e.V. grants to

Endress+Hauser Conducta GmbH & Co. KG

Dieselstrasse 24, 70839 Gerlingen, Germany

the Certificate No: **Z02070** for the PROFIBUS Device:

Model Name: Liquiline CA8xxx
Revision: Profile 3.02; SW/FW: 01.13; HW: 01.00
GSD: EH01155E.GSD, File Version: 20.05.2022
 PA039760.GSD

This certificate confirms that the product has successfully passed the certification tests with the following scope:

<input checked="" type="checkbox"/>	DP-V0	MS0, Freeze, Auto_Baud, Set_Slave_Add
<input checked="" type="checkbox"/>	DP-V1	MS1, Prm_Block_Structure, MS2, I&M
<input checked="" type="checkbox"/>	Profile	PROFIBUS PA 3.02
<input checked="" type="checkbox"/>	Physical Layer	RS485

Test Report Number: 342-1 plus Manufacturer Declaration
Authorized Test Laboratory: ifak, Magdeburg, Germany

The tests were executed in accordance with the following documents:
"Test Specifications for PROFIBUS DP Slaves, Version 3.0.9, September 2008" and
"Test Specifications for PROFIBUS PA Profile 3.0.2, Version 5.0.4, February 2010".

This certificate is granted according to the document:
"Framework for testing and certification of PROFIBUS and PROFINET products".

For all products that are placed in circulation by **January 24, 2026** the certificate is valid for life.

Karlsruhe, June 16, 2025

Board of PROFIBUS Nutzerorganisation e. V.



(Official in Charge)



(Xaver Schmidt)




(Frank Moritz)